

A New Method for Semi-Automatic Volumetry of Cerebral Ventricles in Pediatric Neuroimaging

Horst K. Hahn¹, William S. Millar², Maureen S. Durkin³, Olaf Klinghammer¹, Paola K. Tulipano², and Heinz-Otto Peitgen¹

¹MeVis – Center for Medical Diagnostic Systems and Visualization, Universitaetsallee 29,28359 Bremen, Germany; Email: hahn@mevis.de;

²Diagnostic Neuroradiology Division, Columbia University, 177 Ft. Washington Ave., New York, NY 10032, USA;

³Division of Epidemiology, Joseph L. Mailman School of Public Health, Columbia University, 600 West 168 th St., New York, NY 10032, USA.

Cerebral ventricular volume is an important factor in quantitative neurological diagnosis, for indirectly monitoring the progress of brain damage, e.g. after neonatal white matter damage (WMD) or associated with neurodegenerative diseases, and for assessing postoperative outcome, e.g. on hydrocephalus patients. A prospective, epidemiologic and neuropsychological NIH funded study examines the effects of early brain damage on selective aspects of cognitive development in children comprising the best documentation of localized neonatal brain damage to date. We describe an efficient method to segment and visualize the intracerebral fluid spaces in children and adults based on MRI and to quantify ventricular volumes which is (a) robust (for normal and pathological anatomy), (b) reproducible (less than 2 % variation), and (c) fast (less than 5 min).

Introduction

To investigate the long-term cognitive outcomes of WMD within a prospective, epidemiologic and neuropsychological NIH funded study, a sample of 597 surviving children of very low birth weight (<1501 g) was recruited for thorough neuropsychological evaluations at an age of six years. These children received extensive evaluations at birth, including systematized ultrasonographic (US) brain scans at postnatal days 1-3, 7-10, and 21. A follow-up at age six offers an unprecedented opportunity to test hypotheses about the developing brain and its compensatory response to acquired insults. We hypothesize that cranial magnetic resonance images (MRI) that has been obtained for a sub-sample at age six will show reduced white matter volume and ventricular enlargement associated with presence and location of neonatal WMD documented by US.

None of the tested volumetry methods are suitable to quantify the ventricular volumes in all given 34 given datasets. Fully automated methods failed where anatomical or pathological variability was too large. Manual or two-dimensional methods, such as described in [1] and [2], resulted in long interaction times or poor reproducibility. A volumetry method combining three-dimensional (3d) segmentation based on the watershed transformation [3] with a volumetric histogram analysis [4] produced satisfactory results in cases with medium or strong ventricular enlargement. However, problems occurred with normal pediatric cerebral anatomy that often shows lateral ventricles of 5-10 cm length containing not more than 1 ml of fluid.

Methods

The presented method combines acquisition of standard MRI data, fast 3d marker based segmentation, and automatic histogram analysis. T1-weighted anatomic data is acquired on a GE Signa 1.5 T: GR 3D, 1.5 mm slice thickness, no interslice gap, TR 34 ms, TE 5 ms, matrix 256x256, 124 slices, 0.78 mm pixel spacing.

The segmentation algorithm is a modified version of the fast watershed transformation as described in [3]. Five different marker types are used for ventricle labeling (R, L, 3, and 4) and region exclusion, thereby imposing watersheds at respective borders [4]. Mislabeled regions are interactively attributed to the correct structure by additional markers. The watershed transformation takes approx. 1 sec on a standard PC (Pentium III, 700 MHz, 256 MB) for a typical region of interest (1 million voxels). It automatically tracks the ventricular boundaries in 3d, taking the marker positions into account. In a standard case, 10 markers suffice to define the ventricular anatomy accurately. The segmentation procedure, including user interactions, takes 2 min for all slices on average.

The volumes of each of the four segmented ventricles are computed automatically from corresponding over-inclusive regional image histograms. Assuming symmetric, equally distributed partial volume effects, we use a set of mixed Gaussians to extend the trimodal normal distribution model [5]. Least squared error minimization is used to fit the model. The expected volumetric uncertainties are calculated individually, based on image noise, contrast, and resolution.

The histogram analysis was modified for very narrow ventricles (n = 15 cases) that typically occur in pediatric neuroimaging. In these images, partial volume effects play a more dominant role than we presumed when developing the methods for quantitative neuroimaging on adults. The histogram analysis then automatically rejects the model fit and requests the user to specify at least 10 anatomical positions that are completely surrounded by cerebro spinal fluid (CSF); their mean gray value is given to the histogram model as center of the CSF Gaussian.

Results

With the above modification, also the smallest ventricles were reliably quantified in all 34 datasets. We thoroughly evaluated intra- and inter-observer variance. Mean values and standard deviations have been recorded. Inter-observer variance for lateral ventricular volumes was less than 1 % even for small volumes. Larger relative variations have been observed for 3rd and 4th ventricles due to small object sizes and the frequently imprecise delineations in MRI. Repeated acquisitions on a ventricle shaped paraffin phantom yielded a relative inter-examination variance of 0.4 % for the total volume (Mean ± SD: 60.89 ± 0.22 ml).

Finally, we put emphasis on testing the above modification of the histogram analysis against systematic errors (n = 19 cases). Placing the additional markers resulted in an increased mean interaction time (approx. 3 min), but did not produce biased results. However, a slightly larger inter-observer variance was recorded; i.e., an additional source of error has been created and requires special care of the operator.

Conclusion and Future Perspectives

We present a new semiautomatic approach which allows to quantify pediatric ventricular anatomy accurately while reducing image analysis and interaction times compared to manual or semiautomatic slice-based evaluation. User-induced errors are minimized by placing markers inside the objects instead of tracing object borders interactively. The 3d segmentation procedure is likewise applicable to normal and pathological anatomy since it does not require any anatomical model. Furthermore, asymmetry of the lateral ventricles is directly quantified.

Accurate measurements are achieved from commonly available high-resolution T1-weighted MR images. Image fusion and higher dimensional image analysis were avoided. A model based histogram analysis robustly accounts for image noise, non-uniformity, and partial volume effects. Combining short interaction times, applicability to pathological anatomy, and high reproducibility, the presented method meets the requirements posed by imaging and workflow conditions in a clinical setting.

At the moment of print, the correlation of the volumetry results with the neonatal data was not yet completed, but will be published as soon as available. Recently, the presented method has been successfully applied to MR data of a preterm baby two weeks after birth. Employing a modified histogram model, also the brain volume of the baby has been quantified.




Fig. 1. segmentation result for two 6yo subjects, volumes in ml. a-c: normal neonatal US (L: 1.72±0.20; R: 1.74±0.22; 3: 0.32±0.07; 4: 1.58±0.05). d-f: WMD, predominantly on the left hemisphere (L: 47.47±0.22; R: 9.52±0.17; 3: 2.51±0.07; 4: 1.23±0.06).

References

1. Tsunoda, A., Mitsuoka, H., Sato, K., et al., *Neuroradiology* 42: 424-429, 2000.
2. Schierlitz, L., Hüppi, P.S., Jakob, M., et al., *ISMRM 2001, Proc. Intl. Soc. Mag. Reson. Med* 9: 402, 2001.
3. Hahn, H.K., Peitgen, H.-O., *MICCAI 2000, LNCS 1935: 134-143*, Springer, Berlin, 2000.
4. Hahn, H.K., Lentschig, M.G., Deimling, M., Terwey, B., Peitgen, H.-O., *CARS 2001, ICS 1230: 384-389*, Elsevier, Amsterdam, 2001.
5. Hahn, H.K., Peitgen, H.-O., *3rd Caesarium – Computer Aided Medicine*, Bonn: Springer, Berlin, 2001, in press.