

On the Feasibility of White Matter Arterial Spin Labeling Measurements

P. van Gelderen¹, J. A. de Zwart¹, and J. H. Duyn¹

¹Advanced MRI, LFMI, NINDS, National Institutes of Health, Bethesda, MD, United States

Introduction

Cerebral perfusion can be used to distinguish between normal and compromised brain function, and its measurement has many important applications in the study of brain function and pathology. MRI-based arterial spin labeling (ASL) techniques have been used (1-3) to measure perfusion successfully in various applications (e.g. (4-7)). ASL techniques have a number of drawbacks, such as limited coverage, low SNR and sensitivity to the transit time. An additional complication is the inherently high contrast between gray matter (GM) and white matter (WM), as WM has an approximately 5 fold lower perfusion. This makes the typically low-resolution perfusion measurements sensitive to partial volume and blurring effects. As a result GM signal is mixed into the WM, artificially increasing the measured WM perfusion. On the other hand the longer transit times in WM tend to decrease the WM signal. Both effects complicate the measurement of WM perfusion, which may explain why there is little consistency in the reported MR literature values (8) or with the assumed gold standard of autoradiography (9). Here we discuss the SNR, image resolution and transit time issues in order to estimate WM perfusion reliability. Supporting 3T data is shown as well.

Image Resolution

The limited image resolution results in mixing of the much stronger GM signal into the WM. This can be a result of both partial volume effects, since a voxel may contain a mix of tissues, and of blurring, where signals originating in one voxel are mixed into their neighboring (or farther) voxels by the imaging process. A simulation of these effects is shown in Table 1, demonstrating the contamination of WM signal by the neighboring GM. To limit this effect to less than 50% of the WM signal, which is about 0.2 times the GM signal, a resolution on the order of 1.9mm is necessary.

Transit time

To estimate the WM transit times, bolus tracking experiments were performed on six volunteers in a 3T GE scanner with a 16-channel receive coil and a receiver system build in-house. A twelve-slice EPI series was acquired at 1.5mm resolution, with SENSE rate 2, 1s TR, 40ms TE, 180 repetitions. The fitting of gamma-variate curves resulted in a voxel-by-voxel arrival time of the bolus. Although this timing information does not result in absolute ASL transit times, it does indicate the relative voxel to voxel differences. The spread in bolus arrival was several seconds, with WM mostly showing a later arrival, see Fig. 1. This shows that the transit time, in at least part of the WM, is often several seconds longer than the average GM transit time.

ASL SNR estimation

Assuming a 100% labeling efficiency and complete suppression of the blood signal, the perfusion SNR in a pulsed ASL experiment can be expressed as:

$$SNR_{p,max} = T_1 \cdot f / \lambda \cdot e^{-1} \cdot e^{-t/T_1} \cdot SNR_{base} \cdot \sqrt{N} \cdot 2 \cdot \sqrt{0.5}$$

The following parameters were used: $T_1 = 1500ms$, f_{WM} (perfusion fraction)= 0.003ml/g/s (=0.18ml/g/min), $f_{GM} = 0.013ml/g/s$ (= 0.8ml/g/min) (10, 11), λ_{WM} (partition coefficient) = 0.82ml/g, $\lambda_{GM} = 0.98ml/g$ (12), N (averages)= 92 (i.e. a 10min scan time), $SNR_{base,WM} = 13$, $SNR_{base,GM} = 22$ per unit volume (values measured with a 16 channel array coil at 3T). The resulting $SNR_{p,max}$ can now be calculated for the desired 1.9mm resolution as function of transit time t_t , see Table 2. While the perfusion signal in GM at a 0.9% relative amplitude is easily detectable, the WM amplitude is much lower, even for a short transit time of 1s. Longer transit times result in more label decay, diminishing the relative amplitude to below 0.1%, which is approximately the noise level for this type of experiment, even with the use of a sensitive 16-channel receive array.

ASL experiments

FAIR (3) ASL measurements were performed on 7 volunteers on the same 3T scanner (see above). A 1.9mm resolution EPI was chosen with a TE of 27ms, TR of 3.2s, 10min. scan time. Three scans were taken with labeling delays of 1.5, 2 and 3s. The average perfusion signal in GM and deep WM was calculated using anatomically defined masks. Table 3 reports the resulting SNR and relative amplitudes, showing good correspondence with the theory. The WM perfusion SNR is less than one for all labeling times.

Discussion

FAIR ASL SNR in WM is insufficient for a reliable and practical WM perfusion measurement with the current state of the technology. The WM perfusion fraction is low compared to GM and the transit times are long, probably up to several seconds longer than GM, resulting in a very low perfusion signal in at least part of the WM. Increasing the SNR by increasing the voxel size is not a viable option because it results in excessive contamination from the much stronger GM signals. A further complication is that the transit times appear quite heterogeneous, so that a correction would require a reliable transit time measurement at a similarly high resolution, running into the same SNR limitations.

References

1. J. A. Detre, et al., *Magn Reson Med* **23**, 37 1992.
2. K. K. Kwong et al., *Proc Natl Acad Sci U S A* **89**, 5675 1992.
3. S. G. Kim, *Magn Reson Med* **34**, 293 1995.
4. X. Golay, J. Hendrikse, J. Van Der Grond, *J Neuroradiol* **32**, 321 2005.
5. J. Kim et al., *Neuroimage* **31**, 376 2006.
6. J. Wang et al., *Proc Natl Acad Sci U S A* **102**, 17804 2005.
7. J. A. Chalela et al., *Stroke* **31**, 680 2000.
8. T. R. Steger, R. A. White, E. F. Jackson, *Magn Reson Med* **53**, 895 2005.
9. M. Reivich, J. Jehle, L. Sokoloff, S. S. Kety, *J Appl Physiol* **27**, 296 1969.
10. I. Law et al., *J Cereb Blood Flow Metab* **20**, 1252 2000.
11. B. K. Oowler et al., *J Cereb Blood Flow Metab* **24**, 17 2004.
12. P. Herscovitch, M. E. Raichle, *J Cereb Blood Flow Metab* **5**, 65 1985.

Table 1. Average GM and WM signals after blurring a simulated 1mm resolution perfusion map (based on a 3D anatomical scan) with the GM signal set to one and WM to zero.

Res. [mm]	1.1	1.9	2.3	3.5
GM	0.86	0.78	0.72	0.63
WM	0.05	0.08	0.11	0.17

Figure 1. Distribution of arrival times in a bolus tracking scan, reflecting the start of the signal decrease. The shading indicates the fastest and slowest 10% of the WM.




Table 2. Theoretical relative perfusion amplitude and SNR in GM and WM for a 10min. 1.9mm resolution FAIR ASL scan.

Tissue, t_t [s]	GM	WM	WM	WM
	0.75	1.0	2.0	3.0
SNR	10.1	1.3	0.7	0.3
amp. [%]	0.91	0.21	0.11	0.05

Table 3. Average relative perfusion amplitude and SNR in GM and WM for a 10min 1.9mm resolution FAIR scan

T_{label}	amplitude [%]			perfusion SNR		
	1.5	2.0	3.0	1.5	2.0	3.0
GM	0.87	0.86	0.53	9.37	9.37	5.74
WM	0.10	0.14	0.11	0.64	0.88	0.72