

Nuts & Bolts of Advanced Imaging: Parallel Transmit Pulse Design

William A Grissom

Biomedical Engineering and Radiology, Vanderbilt University, Nashville, Tennessee, USA
will.grissom@vanderbilt.edu

Highlights

- Small- and large-tip-angle parallel transmit pulse design are both based on the inversion of a forward model for excitation, but require different solution strategies.
- Unlike conventional multidimensional pulse design, B_1^+ magnitude and relative phase maps for each transmit coil are required for parallel transmit pulse design.
- For certain pulse types as spokes, the design problem can be dramatically simplified and accelerated.
- For in-vivo applications, RF power and SAR evaluation and control is crucial.

Target Audience: Students and researchers interested in applications of multidimensional parallel excitation, and in the development of parallel transmit RF pulse design algorithms.

Outcome & Objectives: From the theory lecture of the course, participants will learn about approaches to small- and large-tip-angle parallel transmit pulse design for specific applications of parallel transmission. In the interactive hands-on portion of the course, participants will apply the knowledge gained from the theory lecture to design small- and large-tip-angle multidimensional parallel excitation pulses.

Purpose: Magnetic resonance imaging (MRI) at high field strengths higher promises to enhance clinical diagnosis and enable new medical discoveries. However, at high field inhomogeneous transmit RF (B_1^+) fields cause spatially-varying image contrast and even complete loss of signal, severe B_0 inhomogeneity causes signal loss near regions of high susceptibility differences, and increased specific absorption rate (SAR) poses a serious safety risk. Parallel transmission has the potential to address these problems, directly and indirectly, by providing an additional spatial encoding mechanism (the multiple channels and coils of a parallel transmit system) that enhances one's ability to tailor excitation pulses to specific patients and anatomy. In this course students will learn how to take advantage of those new degrees of freedom in RF pulse design.

Methods: Small-tip-angle multidimensional designs are simplified by linearity of the spatial excitation pattern in the RF pulses, which enables pulse design by matrix inverse or iterative methods. Pulse design requires the co-design of an excitation k-space trajectory, which is based on considerations of pulse duration, excitation FOV, and the required sharpness of the excitation pattern. The B_1^+ patterns of each coil must be included in the forward excitation model, and are generally measured in each subject prior to pulse design. Target excitation patterns and excitation error weighting can be constructed to approximately achieve target flip angle ripple levels. SAR evaluation is straightforward for a given set of pulses, but ideally SAR is constrained during pulse design. Large-tip-angle parallel transmit pulse design is generally based on extensions of optimal control to the multi-channel case. Selection of an excitation trajectory for a large-tip-angle pulse in some cases requires additional considerations.

Discussion and Conclusion: Parallel transmit pulse design is a rich problem space, with significant need both for better general-purpose pulse design algorithms, and for the application and adaptation of existing algorithms to specific applications. This course will be a springboard to help participants get started in developing their own parallel transmit pulse design algorithms.